

City of Belleville Fire Department

Number: 1039

POLICY:

It is the policy of the Belleville Fire Department that the Multi Gas Monitor be utilized at all incidents that require air quality monitoring of the atmosphere.

Date S.O.G. Comes Into Effect: May 9, 2011

Date S.O.G. Revised: November 6, 2018

Date Committee Approved S.O.G.: May 9, 2011

Page 1 of 2

GUIDELINE

1:00 **Purpose:**

1:01 To ensure that all personnel are familiar with the operation of the Multi Gas Monitor.

2:00 **Procedures:**

2:01 The Multi Gas Monitor shall be utilized at calls such as; Carbon Monoxide, Smell of Gas, Oxygen deficient atmosphere or any other incident that requires air quality monitoring.

2:02 The Officer in Charge shall ensure that the Multi Gas Monitor is bump tested daily and the battery charged as required. Bump testing instructions will be located with the Test Station.

THIS GUIDELINE IS NOT EXPECTED TO SUBSTITUTE FOR GOOD JUDGEMENT AND EXPERIENCE UNDER UNUSUAL CONDITIONS.

City of Belleville Fire Department

- 2:03 For further information on the use of this Multi Gas Monitor, please refer to the instruction manual.
- 3:00 **Scope:**
- 3:01 It shall be the responsibility of the Captain/ Officer in charge of each shift to ensure that this guideline is explained and followed.
- 3:02 It shall be the responsibility of **all employees** to understand and adhere to this guideline.

THIS GUIDELINE IS NOT EXPECTED TO SUBSTITUTE FOR GOOD
JUDGEMENT AND EXPERIENCE UNDER UNUSUAL CONDITIONS.